

## CERTIFICATE OF ANALYSIS

---

Lot Number: 610M3401  
PRODUCT V-PVP 7%  
Product Number V-PVP-7  
Expiry Date: 31.10.2017

TEST	SPECIFICATION	RESULT
pH	7,25 - 7,45	7,43
OSMOLALITY	300 - 330 mOsm/kg	303 mOsm/kg
ENDOTOXIN	< 0,5 EU/mL	< 0,05 EU/mL
MEA expanded blastocyste after 96 h	≥ 80 %	PASSED
STERILITY	PASSED	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 02.11.2016

