



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 5239 1976  
Fax +49 3641 539 1977

info@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

Lot Number: 610M3601  
PRODUCT V-HYLASE  
Product Number V-HYA  
Expiry Date: 31.10.2017

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,40	7,39
OSMOLALITY	257-273 mOsm/kg	262 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,0 EU/mL	< 0,2 EU/mL
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 28.10.2016