

# CERTIFICATE OF ANALYSIS

PROVISIONAL

---

**Lot Number:** 611M1401  
**PRODUCT** V-ONESTEP Culture Medium  
**Product Number** V-OSM-20  
**Expiry Date:** 12.02.2017

TEST	SPECIFICATION	RESULT
pH	7,25 - 7,40	7,33
OSMOLALITY	257-273 mOsm/kg	261 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blactocyst in 120h	≥ 80 %	PASSED
Color: Pink rose	PASSED	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 23.11.2016

