



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

info@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 6120112/113/114/115/117/120/130/160
PRODUCT NAME: V-DENUPET 125, 135, 140, 150, 175, 200, 300, 600
REFERENCE: V-DEN-125, -135, -140, -150, -175, -200, -300, -600
EXPIERY DATE: 12.12.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 28.12.2016

i. A. J. Slaba