

## CERTIFICATE OF ANALYSIS

---

Lot Number: 612HO30  
PRODUCT V-ICSIPET HOLDING 30°  
Product Number V-HOLD-30  
Expiry Date: 26.12.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 24.01.2017

