

CERTIFICATE OF ANALYSIS

LOT NUMBER: 612IN302/IN303/IN353/352
PRODUCT NAME: V-ICSIPET INJECTION 30°, 35°
REFERENCE: V-INJ-S3-30,35 / V-INJ-S2-30,35
EXPIERY DATE: 26.12.2019

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 24.01.2017

