

CERTIFICATE OF ANALYSIS

Lot Number: 612M0101
PRODUCT V-GRAD 100%
Product Number V-GRD-100
Expiry Date: 30.06.2018

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,60	7,58
OSMOLALITY	300 - 330 mOsm/kg	316 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 02.01.2017

