

CERTIFICATE OF ANALYSIS

Lot Number: 612M0201
PRODUCT V-GRAD 80%
Product Number V-GRD-80
Expiry Date: 30.06.2018

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,60	7,37
OSMOLALITY	310 - 330 mOsm/kg	316 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,0 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 12.01.2017

