

CERTIFICATE OF ANALYSIS

Lot Number: 612M0301
PRODUCT V-GRAD 40%
Product Number V-GRD-40
Expiry Date: 30.06.2018

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,60	7,38
OSMOLALITY	310 - 340 mOsm/kg	324 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,0 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 12.01.2017

