

CERTIFICATE OF ANALYSIS

Lot Number: 612M0601
PRODUCT V-SPERM FREEZE
Product Number V-SFM-12
Expiry Date: 30.06.2018

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,60	7,41
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 12.01.2017

