

CERTIFICATE OF ANALYSIS

Lot Number: 612M1401
PRODUCT V-ONESTEP Culture Medium
Product Number V-OSM-20
Expiry Date: 12.03.2017

TEST	SPECIFICATION	RESULT
pH	7,25 - 7,40	7,37
OSMOLALITY	257-273 mOsm/kg	266 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blactocyst in 120h	≥ 80 %	PASSED
Color: Pink rose	PASSED	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 27.12.2016

