

CERTIFICATE OF ANALYSIS

Lot Number: 612M3001
PRODUCT V-OIL 100 P
Product Number V-OIL-P100
Expiry Date: 30.06.2018

TEST	SPECIFICATION	RESULT
POV	< 0,1 mEq/Kg	< 0,1 mEq/Kg
ENDOTOXIN	< 0,25 EU/mL	< 0,1 EU/ml
MEA 1-cell to expanded blactocyst in 120h	≥ 80 %	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 16.12.2016

