



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 5239 1976  
Fax +49 3641 539 1977

info@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

Lot Number: 612M3101  
PRODUCT V-OIL 500 P  
Product Number V-OIL-P500  
Expiry Date: 30.06.2018

TEST	SPECIFICATION	RESULT
POV	< 0,1 mEq/Kg	< 0,1 mEq/Kg
ENDOTOXIN	< 0,25 EU/mL	< 0,1 EU/ml
MEA 1-cell to expanded blactocyst in 120h	≥ 80 %	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 23.12.2016