

## CERTIFICATE OF ANALYSIS

---

Lot Number: 7010114  
PRODUCT V-DENUPET 140  
Product Number V-DEN-140  
Expiry Date: 20.01.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 28.02.2017

