

CERTIFICATE OF ANALYSIS

LOT NUMBER: 7010112/113/117/120/127/130/160
PRODUCT NAME: V-DENUPET 125, 135, 175, 200, 275, 300, 600
REFERENCE: V-DEN-125, -135, -175, -200, -275, -300, -600
EXPIRY DATE: 20.01.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 14.02.2017

