

## CERTIFICATE OF ANALYSIS

LOT NUMBER: 7030115/117/120/127/130/160  
PRODUCT NAME: V-DENUPET 150, 175, 200, 275, 300, 600  
REFERENCE: V-DEN-150, -175, -200, -275, -300, -600  
EXPIERY DATE: 02.03.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 22.03.2017

