

CERTIFICATE OF ANALYSIS

Lot Number: 703M0302
PRODUCT V-GRAD 40%
Product Number V-GRD-40
Expiry Date: 30.09.2018

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,60	7,37
OSMOLALITY	310 - 340 mOsm/kg	320 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	< 0,1 EU/mL
SPERM SURVIVAL TEST	PASSED	PASSED

PRODUCT RELEASED BY Jithin Vasanthan ON 26.04.2017

