

CERTIFICATE OF ANALYSIS

Lot Number: 703M3401
PRODUCT V-PVP 7%
Product Number V-PVP-7
Expiry Date: 31.03.2018

TEST	SPECIFICATION	RESULT
pH	7,25 - 7,45	7,43
OSMOLALITY	300 - 330 mOsm/kg	325 mOsm/kg
ENDOTOXIN	< 0,5 EU/mL	< 0,05 EU/mL
MEA expanded blastocyste after 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 28.03.2017

