

## CERTIFICATE OF ANALYSIS

---

**LOT NUMBER:** 7040112/113/114/115/117/127/130/160  
**PRODUCT NAME:** V-DENUPET 125, 135, 140, 150, 175, 275, 300, 600  
**REFERENCE:** V-DEN-125, -135, -140, -150, -175, -275, -300, -600  
**EXPIERY DATE:** 02.04.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 26.04.2017

---

