



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 5239 1976  
Fax +49 3641 539 1977

info@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

**Lot Number:** 704M0301  
**PRODUCT** V-GRAD 40%  
Product Number V-GRD-40  
Expiry Date: 31.10.2018

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,60	7,38
OSMOLALITY	310 - 340 mOsm/kg	319
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	< 0,1 EU/mL
SPERM SURVIVAL TEST	≥ 80 %	PASSED

PRODUCT RELEASED BY Jithin Vasanthan ON 26.04.2017