

CERTIFICATE OF ANALYSIS

Lot Number: 704M0401
PRODUCT V-GRAD Kit
Product Number V-GRD-KIT
Expiry Date: 31.10.2018

TEST	SPECIFICATION	RESULT
pH (V-GRAD 40)	7,20 - 7,60	7,38
pH (V-GRAD 80)	7,20 - 7,60	7,49
OSMOLALITY (V-GRAD 40)	310 - 340 mOsm/kg	319
OSMOLALITY (V-GRAD 80)	310 - 330 mOsm/kg	311
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	< 0,1 EU/mL
SPERM SURVIVAL TEST	≥ 80 %	PASSED

PRODUCT RELEASED BY Jithin Vasanthan ON 26.04.2017

