

## CERTIFICATE OF ANALYSIS

---

Lot Number: 704M0601  
PRODUCT V-SPERM FREEZE  
Product Number V-SFM-12  
Expiry Date: 31.10.2018

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,60	7,53
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 02.05.2017

