



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

info@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

Lot Number: 704M1401
PRODUCT V-ONESTEP Culture Medium
Product Number V-OSM-20
Expiry Date: 02.07.2017

TEST	SPECIFICATION	RESULT
pH	7,25 - 7,40	7,32
OSMOLALITY	257-273 mOsm/kg	261 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blactocyst in 120h	≥ 80%	PASSED

PRODUCT RELEASED BY Jithin Vasanthan ON 18.04.2017
