

CERTIFICATE OF ANALYSIS

Lot Number: 704M2301
PRODUCT V-HEPES PLUS 100
Product Number V-HTF-P1
Expiry Date: 30.04.2018

TEST	SPECIFICATION	RESULT
pH	7,00 - 7,40	7,14
OSMOLALITY	257-273 mOsm/kg	263 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 05.05.2017

