



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

info@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

Lot Number: 705M0401
PRODUCT V-GRAD Kit
Product Number V-GRD-KIT
Expiry Date: 30.11.2018

TEST	SPECIFICATION	RESULT
pH (V-GRAD 40)	7,20 - 7,60	7,48
pH (V-GRAD 80)	7,20 - 7,60	7,49
OSMOLALITY (V-GRAD 40)	310 - 340 mOsm/kg	320 mOsm/kg
OSMOLALITY (V-GRAD 80)	310 - 330 mOsm/kg	311 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	< 0,1 EU/mL
SPERM SURVIVAL TEST	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 22.05.2017