



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

info@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

Lot Number: 705M0601
PRODUCT V-SPERM FREEZE
Product Number V-SFM-12
Expiry Date: 30.11.2018

| TEST | SPECIFICATION | RESULT |
|---------------------|---------------|--------------|
| pH | 7,20 - 7,60 | 7,50 |
| BACTERIAL SCREEN | PASSED | PASSED |
| FUNGAL SCREEN | PASSED | PASSED |
| SPERM SURVIVAL TEST | ≥ 80 % | PASSED |
| ENDOTOXIN | < 0,25 EU/mL | < 0,05 EU/mL |

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 02.06.2017