

## CERTIFICATE OF ANALYSIS

---

Lot Number: 705M1401  
PRODUCT V-ONESTEP Culture Medium  
Product Number V-OSM-20  
Expiry Date: 31.07.2017

TEST	SPECIFICATION	RESULT
pH	7,25 - 7,40	7,32
OSMOLALITY	257-273 mOsm/kg	264 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blastocyst in 120h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 22.05.2017

