

CERTIFICATE OF ANALYSIS

Lot Number: 706M0401
PRODUCT V-GRAD Kit
Product Number V-GRD-KIT
Expiry Date: 31.12.2018

TEST	SPECIFICATION	RESULT
pH (V-GRAD 40)	7,20 - 7,60	7,43
pH (V-GRAD 80)	7,20 - 7,60	7,41
OSMOLALITY (V-GRAD 40)	310 - 340 mOsm/kg	317 mOsm/kg
OSMOLALITY (V-GRAD 80)	310 - 330 mOsm/kg	320 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 26.06.2017

