



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 5239 1976  
Fax +49 3641 539 1977

info@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

Lot Number: 706M0501  
PRODUCT V-SPERM WASH  
Product Number V-SWM-100  
Expiry Date: 30.06.2018

TEST	SPECIFICATION	RESULT
pH	7,00 - 7,40	7,17
OSMOLALITY	257-273 mOsm/kg	263 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 17.07.2017