



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

info@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

Lot Number: 706M0601
PRODUCT V-SPERM FREEZE
Product Number V-SFM-12
Expiry Date: 23.12.2018

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,60	7,50
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 29.06.2017