

CERTIFICATE OF ANALYSIS

LOT NUMBER: 7070115/117/120/127
PRODUCT NAME: V-DENUPET 150, 175, 200, 275
REFERENCE: V-DEN-150, -175, -200, -275
EXPIERY DATE: 02.07.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 01.08.2017

