

## CERTIFICATE OF ANALYSIS

---

**LOT NUMBER:** 707IN302/IN303/IN352/IN353/IN203  
**PRODUCT NAME:** V-ICSIPET INJECTION 20°, 30°, 35°  
**REFERENCE:** V-INJ-S2-30/35, INJ-S3-20/30/35  
**EXPIRY DATE:** 27.07.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 12.09.2017

