

CERTIFICATE OF ANALYSIS

Lot Number: 707M0401
 PRODUCT V-GRAD Kit
 Product Number V-GRD-KIT
 Expiry Date: 31.01.2019

TEST	SPECIFICATION	RESULT
pH (V-GRAD 40)	7,20 - 7,60	7,43
pH (V-GRAD 80)	7,20 - 7,60	7,37
OSMOLALITY (V-GRAD 40)	310 - 340 mOsm/kg	322 mOsm/kg
OSMOLALITY (V-GRAD 80)	310 - 330 mOsm/kg	320 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 31.07.2017

