

CERTIFICATE OF ANALYSIS

Lot Number: 707M2301
PRODUCT: V-HEPES PLUS 100
Product Number: V-HTF-P1
Expiry Date: 31.07.2018

TEST	SPECIFICATION	RESULT
pH	7,00 - 7,40	7,14
OSMOLALITY	257-273 mOsm/kg	269 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 28.07.2017

