

CERTIFICATE OF ANALYSIS

LOT NUMBER: 7080114/117/127/130
PRODUCT NAME: V-DENUPET 140, -175, -275, -300
REFERENCE: V-DEN-140, -175, -275, -300
EXPIERY DATE: 07.08.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 12.09.2017

