

CERTIFICATE OF ANALYSIS

Lot Number: 708GP15
PRODUCT V-GLASSPET 5.3/4 (150)
Product Number V-GPT-150
Expiry Date: 31.08.2022

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 06.11.2017

