

## CERTIFICATE OF ANALYSIS

---

Lot Number: 708GP23  
PRODUCT V-GLASSPET 9 (230)  
Product Number V-GPT-230  
Expiry Date: 31.08.2022

TEST	SPECIFICATION	RESULT
Gamma Sterilization	< 0,05 EU/mL	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 06.11.2017

