

CERTIFICATE OF ANALYSIS

Lot Number: 708M0601
PRODUCT V-SPERM FREEZE
Product Number V-SFM-12
Expiry Date: 28.02.2019

TEST	SPECIFICATION	RESULT
pH	7,20 - 7,60	7,56
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL

PRODUCT RELEASED BY Jithin Vasanthan ON 07.09.2017

i.A. 