

CERTIFICATE OF ANALYSIS

Lot Number: 708M0901
PRODUCT V-SPERM PREPARATION KIT
Product Number V-SP-KIT
Expiry Date: 31.08.2018

TEST	SPECIFICATION	RESULT
pH	7,00 - 7,40	7,15
OSMOLALITY	257-273 mOsm/kg	263 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blactocyst in 96 h	≥ 80%	PASSED

PRODUCT RELEASED BY ON 30.08.2017

i.A. 