

## CERTIFICATE OF ANALYSIS

---

**Lot Number:** 708M1401  
**PRODUCT** V-ONESTEP Culture Medium  
Product Number V-OSM-20  
Expiry Date: 08.11.2017

TEST	SPECIFICATION	RESULT
pH	7,25 - 7,40	7,28
OSMOLALITY	257-273 mOsm/kg	259 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
MEA 1-cell to expanded blactocyst in 120h	≥ 80%	PASSED

PRODUCT RELEASED BY Jithin Vasanthan ON 25.08.2017

i.A. 