

## CERTIFICATE OF ANALYSIS

---

Lot Number: 708M3201  
PRODUCT V-OIL 100 G  
Product Number V-OIL-G100  
Expiry Date: 28.02.2019

TEST	SPECIFICATION	RESULT
POV	< 0,1 mEq/Kg	< 0,1 mEq/Kg
ENDOTOXIN	< 0,25 EU/ml	< 0,1 EU/ml
MEA 1-cell to expanded blactocyst in 120 h	≥ 80 %	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED

---

PRODUCT RELEASED BY Jithin Vasanthan ON 31.08.2017



i.f.