



**VITROMED**  
Germany

VITROMED GmbH  
Wildenbruchstr. 15  
07745 Jena  
GERMANY

Tel +49 3641 5239 1976  
Fax +49 3641 539 1977

info@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

**Lot Number:** 708M3401  
**PRODUCT** V-PVP 7%  
Product Number V-PVP-7  
Expiry Date: 31.08.2018

TEST	SPECIFICATION	RESULT
ENDOTOXIN	< 0,50 EU/mL	0,0755 EU/mL
OSMOLALITY	300 - 330 mOsm/kg	305 mOsm/kg
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED
pH (37°C, ambient air)	7,25-7,45	7,35
STERILITY	PASSED	PASSED

PRODUCT RELEASED BY Jithin Vasanthan ON 07.09.2017

i.A