

## CERTIFICATE OF ANALYSIS

---

**LOT NUMBER:** 7090113/114/117/120/127/130  
**PRODUCT NAME:** V-DENUPET 135, 140, 175, 200, 275, 300  
**REFERENCE:** V-DEN-135, -140, -175, -200, -275, -300  
**EXPIERY DATE:** 15.09.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 04.10.2017

