

CERTIFICATE OF ANALYSIS

LOT NUMBER: 7100112/113/114/115/117/120/127/130
PRODUCT NAME: V-DENUPET 125, 135, 140, 150, 175, 200, 275, 300
REFERENCE: V-DEN-125, -135, -140, -150, -175, -200, -275, -300
EXPIRY DATE: 20.10.2020

TEST	SPECIFICATION	RESULT
Gamma Sterilization	25 - 45 kGy	PASSED
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blactocyst in 96 h	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 05.12.2017

