

CERTIFICATE OF ANALYSIS

Lot Number: 710HO30
PRODUCT V-ICSIPET HOLDING 30°
Product Number V-HOLD-30
Expiry Date: 14.10.2020

| TEST | SPECIFICATION | RESULT |
|-------------------------------------------|---------------|--------|
| Gamma Sterilization | 25 - 45 kGy | PASSED |
| ENDOTOXIN | < 0,05 EU/mL | PASSED |
| MEA 1-cell to expanded blactocyst in 96 h | ≥ 80 % | PASSED |

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 17.11.2017

