

CERTIFICATE OF ANALYSIS

Lot Number: 710M1401
 PRODUCT V-ONESTEP Culture Medium
 Product Number V-OSM-20
 Expiry Date: 31.12.2017

TEST	SPECIFICATION	RESULT
pH	7,25 - 7,40	7,32
OSMOLALITY	257-273 mOsm/kg	262 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	< 0,05 EU/mL
SPERM SURVIVAL TEST	≥ 80 %	PASSED
Color Pink rose	PASSED	PASSED
MEA 1-cell to expanded blastocyst in 120 h + additional cell count on day 6	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 18.10.2017

