

CERTIFICATE OF ANALYSIS

Lot Number: 710M1402
PRODUCT: V-ONESTEP Culture Medium
Product Number: V-OSM-20
Expiry Date: 23.01.2018

TEST	SPECIFICATION	RESULT
pH	7,25 - 7,40	7,38
OSMOLALITY	257-273 mOsm/kg	263 mOsm/kg
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
Color Pink rose	PASSED	PASSED
MEA 1-cell to expanded blactocyst in 120 h + additional cell count on day 6	PASSED	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 17.11.2017

