

CERTIFICATE OF ANALYSIS

Lot Number: 710M3001
PRODUCT V-OIL 100 P
Product Number V-OIL-P100
Expiry Date: 30.04.2019

TEST	SPECIFICATION	RESULT
POV	< 0,1 mEq/Kg	< 0,1 mEq/Kg
ENDOTOXIN	< 0,25 EU/mL	< 0,1 EU/ml
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED
MEA 1-cell to expanded blactocyst in 120 h + additional cell count on day 6	≥ 80 %	PASSED

PRODUCT RELEASED BY Dr.rer.nat. Franziska Buch ON 20.11.2017

