



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 801M3101
PRODUCT NAME: V-OIL 500 P CE
REFERENCE: V-OIL-P500
EXPIRY DATE: 31.07.2019

| TEST | SPECIFICATION | RESULT |
|--|---------------|--------|
| Peroxide value (POV) determination | < 0,1 mEq/kg | PASSED |
| ENDOTOXIN | PASSED | PASSED |
| STERILITY by sterile filtration over 0,2 µm filter | PASSED | PASSED |
| Blastocyst Cell Count | PASSED | PASSED |
| Time Lapse Mouse Embryo Assay | ≥ 80 % | PASSED |
| 1-cell to expanded blastocyst in 120 h | | |