



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 5239 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 801M3601
PRODUCT NAME: V-HYLASE
REFERENCE: V-HYA
EXPIRY DATE: 31.01.2019

| TEST | SPECIFICATION | RESULT |
|---|-----------------|--------|
| pH | 7,20-7,40 | 7,36 |
| OSMOLALITY | 257-273 mOsm/kg | 269 |
| BACTERIAL SCREEN | PASSED | PASSED |
| FUNGAL SCREEN | PASSED | PASSED |
| ENDOTOXIN | < 1,00 EU/mL | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 % | PASSED |